SALISBURY ODSTOCK CUSHION USER MANUAL STM/SO

The packaging of this cushion is sealed prior to despatch and the cushion itself will be clean prior to closure. If the cushion has become exposed or damaged during transit, please contact us at the address below

CMHR foam contoured cushion. Two way stretch, vapour permeable, removable cover, sewn seams, double sided flap to protect zipped closur to three sides.



FOAM FOR CARE
Azalea Close, Clover Nook Industrial Park,
Alfreton, Derbyshire DE55 4QX
Tel 01773 838800 Fax 01773 838870

The manufacturing process operates under Class 1 of the Medical Devices Directive by fully complying with the essential requirements for CE marking this product.

Flammability

The Salisbury Odstock Cushion meets the requirements of BS6807:2006 for medium hazard.

The interior materials meet the requirements of The Furniture & Furnishings (Fire) (Safety) Regulations 1988.

User Instructions

This Salisbury Odstock Cushion has been designed for use in wheelchairs and/or day care chairs, by patients who require additional comfort♣, with a weight guide of up to 127kg / 20 Stone.

Ensure that the surface of the chair seat is clear from debris before placing the cushion as debris can damage the coating of the fabric.

The label on this cushion has a unique batch number which should be used in the event of a warranty claim. NEVER remove the label.

The cushion should be placed in the seat of the chair with the castellated surface uppermost. Instructions for use are screen printed onto the cover of the product for each of reference.

♣ *Indication of use based upon interface pressure readings– not a prescriptive criteria. Always defer to clinician's judgement. Safe working load based on test methods devised at the Recticel International Development Centre, Wetteren, Belgium.*

Warranty

Covered against faulty manufacture and materials as replacements goods.

Normal wear and tear expected would not be included.

All warranties subject to terms and conditions of trading.

Please refer to our warranty schedule for terms.

In order to claim under warranty refer to the batch number printed on the product label.

Removing the Cover.

The cover of this cushion is removable for cleaning purposes. Always ensure that, if removed, the cover is returned to the same foam core. Using this cover on a different cushion core MAY render the warranty invalid.

Always sit on the plum side of the cover NOT the grey. The profiled foam is the upper surface and should be placed into the lid of the cover with the flat surface on the base.

The flap covers a water-resistant zip on three sides. This zip should be positioned to the sides and rear of the cushions.

Following the above instruction will ensure that there is no front sea, other than that at the lower edge of the cushion.

Instructions for use are screen printed onto the base of the product for ease of reference.

Cleaning

For general cleaning and superficial dirt, use a disposable wipe with a solution of warm water and neutral detergent. Do not use abrasive or phenolic-based cleaners.

To disinfect in situ, use a 0.1% sodium hypochloride solution (1,000 ppm chlorine). If required, a 1% chlorine solution (10,000ppm) and cold water may be applied for not more than 2 minutes

Thoroughly rinse and dry the cover before re-use or storage. Store in a cool dry area.

Removable contaminated covers should be laundered at 65°C for not less than 10 minutes or at 95°C for not less than 3 minutes.

Do not mangle or iron.

Storage & Care

- Store the cushion flat – do not fold
- Do not place heavy objects on the cushion surface when not in use
- Do not allow sharp objects to penetrate the cover coating
- Keep the cushion clean and use protective cover when not in use
- Regularly inspect the exterior of the cushion
- Do not store in damp conditions or next to radiators or other heating devices.
- Do not store directly onto the floor